PREVENTIVE MAINTENANCE CHECKLIST
SIX (6) MONTH

SERIAL NO. _____ **DATE** _____

MACHINE I.D. NO. _____ **MACHINE HOURS** _____

TECHNICIAN(S) _____ **LABOR HOURS** _____

See the Referenced Section of the 2008K^2 Preventive Maintenance Procedures (part number 508138 Rev. I) for detailed instructions to perform the following:

Reference Section	Procedure	Value	Completed By:
2.1	**Clean Filters**.....(Replace If Necessary)		
2.2	**Inspect Pre-UF Pump filter**.....(Replace If Necessary)		
2.3	**Inspect UF Pump Check Valves**.....(Replace If Necessary)		
2.4	**Check High Voltage AC Connections**.....(≤0.2 ohms)		
2.5	**Calibrate UF Pump**.....(Range: 24 ± 0.1cc at 24 UF Pump strokes)		
2.6	**Check Conductivity Calibration**.....(Range: Δ ± 0.1mS/cm)	Machine Display: _____	External Meter: _____
2.7	**Check Temperature Calibration**.....(Range: 37 ± 0.3°C)		
	(Range: 39 ± 0.3°C)		
2.8	**Check Alarm Operation and Pressure Holding Test**.....		
2.9	**Verify pH**.....(Range 6.9 to 7.6)		
2.10	**Check Power Failure Alarm**.....		
2.11	**Final Checks**.....		

NOTES:

_____

_____

_____

_____

_____

_____

_____

_____

Print Form